

ASPEN SIERRA™


LSO 637 | TLSO 464

Doctor: _____ Fitter: _____

Patient Name: _____ Date: _____

Patient #: _____ Additional Follow-Up Dates: _____

TOOLS NECESSARY: Scissors • Heat Gun • Tape Measure



SIZING THE ASPEN SIERRA

DETERMINE CIRCUMFERENCE

Average Lower Rib and Hip Circumference.

Lower rib circumference = _____
 Hip circumference = _____
 Average = _____

SELECT CONFIGURATION

Select option based on Average


<input type="checkbox"/> 28-33 in 71-84 cm	S
<input type="checkbox"/> 32-44 in 81-112 cm	S + M
<input type="checkbox"/> 43-51 in 109-130 cm	S + L
<input type="checkbox"/> 50-59 in 127-150 cm	S + L + M

DETERMINE PLACEMENT OF COMPRESSION MECHANISM

Average - 40 in = _____ (A)


If (A) is a negative number the mechanism will overlap on the back panel.

$(A) \div 2 = \text{_____} (B)$



If (A) is a positive number, (A) will equal the distance between each mechanism.

Distance Between Mechanism = (A)



ASPEN SIERRA™

LSO 637 | TLSO 464


Doctor: _____ Fitter: _____

Patient Name: _____ Date: _____


Patient #: _____ Additional Follow-Up Dates: _____

TOOLS NECESSARY: Scissors • Heat Gun • Tape Measure


SHORT ANTERIOR PANEL




TALL ANTERIOR PANEL



LSO BACK PANEL



TLSO BACK PANEL





SHORT T-BAR/SHORT ANTERIOR PANEL


REG T-BAR/SHORT ANTERIOR PANEL

SHORT T-BAR/TALL ANTERIOR PANEL

REG T-BAR/TALL ANTERIOR PANEL



STANDARD



STANDARD

